

SUPERVISOR TEST


LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/07/2024

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:23
INTERNAL STANDARD	VERIFIED	03:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:24
BLANK TEST	0.00	03:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:25
BLANK TEST	0.00	03:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:26
BLANK TEST	0.00	03:27



SUPERVISOR TEST



LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/15/2024

OPERATOR NAME:
MAKENZIE A HAYES
Permit #: 49532
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 24120

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:33
INTERNAL STANDARD	VERIFIED	01:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:33
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:34
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:35
BLANK TEST	0.00	01:36



SUPERVISOR TEST




LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/21/2024

OPERATOR NAME:
CHAD P BLINN
Permit #: 33034
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 24140

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:55
INTERNAL STANDARD	VERIFIED	03:55
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:55
BLANK TEST	0.00	03:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:56
BLANK TEST	0.00	03:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:58
BLANK TEST	0.00	03:58



SOLUTION CHANGE

LAKE PLACID VILLAGE
POLICE DEPT.
DATAMASTER dmt: 100290

Date: 07/28/2024

OPERATOR NAME:
MAKENZIE A HAYES
Permit #: 49532
Operator Agency/Dept: LAKE PLACID VILL. PD
Reference Standard #: 24190

Nominal = 0.10 g/210L

BLANK TEST	0.00	19:28
INTERNAL STANDARD	VERIFIED	19:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:28
BLANK TEST	0.00	19:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:30
BLANK TEST	0.00	19:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	19:31
BLANK TEST	0.00	19:31

